

40 75 21.23 – Process Liquid Analytical Measurement (Free-Chlorine Analysers)

COLORIMETRIC FREE CHLORINE ANALYSER

PART 1- GENERAL

SUMMARY

A programmable Free Chlorine Controller/Analyser System shall be supplied for monitoring and/or control of free chlorine in water.

AMPEROMETRIC FREE CHLORINE ANALYSER

1. Controller/Analyser Requirements

- a. Provide monitoring and control of free residual chlorine.
- b. Have capability to control the dosing of free chlorine, including flow-adjusted dosing.
- c. Allow the setup of multiple password-enabled user logins, with access logs available. Logins will have configurable levels of access to system settings.
- d. Provide viewable and downloadable data logs, status logs and service reports.
- e. Provide options for the communication of data, statuses and errors to local SCADA systems (or similar).

2. Standards and Certifications

- a. The manufacturer shall provide a controller enclosure rated to IP65/NEMA 4X. Any accompanying enclosures, e.g. for the purpose of expanding the capacity of the controller, shall also be rated to IP65/NEMA 4X.
- b. The controller system shall be manufactured in accordance with ISO 9001:2015, in an externally audited quality management system.

SYSTEM FEATURES

1. Control Feature Requirements

The controller will be capable of (either by default or optionally):

- a. Providing dosing outputs in the following operator selectable modes: off, manual, automatic
- b. Dosing via PID (Proportional/Integral/Derivative) algorithms. The controller will be able to provide a percentage output to chemical feed pumps for true PID control via 4-20mA signal and/or relay (digital output) in on/off, pulse width or pulse frequency modes
- c. PID control which will have the following options and/or safety features:
 - i. Dedicated failure mode, where dosing can be set to 'off' or a pre-defined percentage value in the event of an error state
 - ii. Be capable of accepting a site stop signal to cease dosing
 - iii. Be able to limit dosing % output to defined minimum and maximum values
 - iv. Configurable overfeed protection, in the event of a failure of dosing hardware (e.g. blocked injectors, empty chemical tanks and the like)
 - v. Viewable PID diagnostic screen, illustrating the extent to which the elements of the algorithm (proportional/integral/derivative) are contributing to the total percentage output
- d. Responding to the loss of sample flow (detected via integrated flow switch) by entering an alarm state and switching active PID dosing to off/a set manual % value
- e. Programmable events can be set via timer (e.g. every 24 hours) or scheduler (e.g. day/week/month)
- f. Providing viewable and downloadable data logs, status logs and service reports on controller operations

2. Hardware Requirements

The controller shall be supplied with:

- a. A colorimetric free chlorine sensor, incorporating:
- b. A suitable flow switch for the detection of sufficient sample water flow

Additionally, the following options/upgrades must be made available:

- a. Additional and/or replacement sensors (including, but not limited to):
 - i. Total chlorine
 - ii. pH
 - iii. ORP/Redox
 - iv. Conductivity
 - v. Alkalinity
 - vi. Turbidity
- b. Digital outputs (relays) up to a maximum of 8 per enclosure
- c. Digital inputs up to a maximum of 8 per enclosure
- d. Analogue outputs (4-20mA) up to a maximum of 4 per enclosure
- e. A backboard incorporating both controller/analyser and flow cell assembly

The system should also be expandable, by the addition of extra enclosures (or similar) to increase the capacity of an individual system beyond the limits stated above.

3. Communication Requirements

In addition to the outputs stated above, the controller shall have the followings features available either by default or optionally:

- a. Digital industrial protocols, including; Modbus RS485, Modbus TCP/IP, Profibus DPV
- b. LAN Ethernet or GPRS connectivity, providing access to an MQTT server system that allows two-way remote communication with the controller/analyser. This communication should include the ability to remotely alter systems settings and download service/access logs, although this feature should be limited by the security/access requirements below.
- c. The controller shall be accessible with a standard Internet browser using a fully interactive graphical interface that can include multiple accounts of varying security and access levels. The controller shall be capable of providing real-time monitoring and control via a personal computer, smartphone, or tablet device. Controllers that require software or apps for communications shall not be considered equal.
- d. The controller shall be capable of e-mail and SMS text alerts upon alarm activation. Controllers without the option of email and text notifications shall not be considered equal.
- e. Online data logging/storage for 30 days or more, with this data being made available for download (in an easily accessible, non-proprietary file format) via Internet browser.
- f. The controller shall include at least two programmable threshold levels and at least two programmable alarm levels for each sensor parameter. Both audible and visible alarms should be provided with the controller/analyser.

PART 2- PRODUCTS

SENSORS

1. Sensor Requirements

The sensor for the monitoring of free chlorine levels shall utilize the following technology:

- a. A method of colorimetric analysis that is not affected by changes in sample pH.
- b. Although optional, does not require calibration in situ.
- c. Incorporates a flow switch, for the detection of the loss of sample flow.
- d. Performs a colorimetric test for chlorine every 2.5 minutes.
- e. Uses an LED emitter with a peak wavelength of 515nm.
- f. Performs a blank 'zero' measurement prior to each colorimetric test, so account for changes in turbidity, colour etc.
- g. Can operated continually for one month without operator intervention.
- h. Is capable of displaying reagent levels on the screen of the controller/analyser.
- i. Has maintenance notifications including reagent replacement and tubing replacement.

2. Sensor Specifications

a. Free Chlorine:

Type: DPD colorimetric

Range: 0-5 mg/l

Sample frequency: 2.5 – 100 minutes

Limit of detection: 0.03 mg/l

Resolution: 0.01 mg/l

Stability: approx. +/-1% or less per month

Wavelength: 515nm

Temperature range: 0-45°C

pH range: pH 4-12

Zero point adjustment: Not necessary

Calibration: None

WARRANTY

The controller electronics and sensors shall be covered by a standard manufacturer warranty of at least one (1) year.

PART 3 EXECUTION

Preparation

The flow cell assembly shall be provided pre-mounted onto a backboard, to allow for easy installation onto a suitable wall

Storage temperature should be 0 to 60 degrees Celsius

Contractor will install the transmitter and sensors in strict accordance with the manufacturer's instructions and recommendation, including:

- a. Suitable continual electrical supply (100-240VAC, or 9-36VDC in the case of DC-specified systems)
- b. Suitable continual flow of sample water (approx.. 500ml/min)